

TD2019IRMS: Summary of Major Modifications

The Technical Document on Detection of Synthetic Forms of Endogenous Anabolic Androgenic Steroids by GC-C-IRMS, TD2019IRMS, has been modified to update the formula for adjustment of concentrations in *Samples* with high (> 1.018) urinary specific gravity (SG).

The formula in footnote 3 has been updated to reflect the changes made in the new TD2019DL.

In addition, the requirement to include the confirmed concentrations of the *Markers* of the steroid profile in the Test Report for IRMS has been removed, since the reporting of such values is mandatory within the steroid profile module in *ADAMS*.

The TD2019IRMS replaces the former TD2016IRMS.